

INDIAN ASSOCIATION FOR THE CULTIVATION OF SCIENCE

2A & 2B , Raja S.C. Mullick Road, Jadavpur, Kolkata-700032

NIT No. : IACS/SPS/NP/2018-19/91

Date: 23/01/2019

Sealed tender in two bids system (technical bid and price bid) is invited from bonafide, resourceful and eligible manufacturer/exclusive distributor/vendors for **Mobile Optical Spectrophoto/Fluorometer**

Part-I (Technical Bid) of the tender should contain technical specifications in detail as well as commercial terms and conditions. Part-II (Price Bid) should clearly indicate group-wise price, if needed, as mentioned in the technical bid. The Technical Bid and Price Bid are to be submitted in separately sealed envelopes, distinctly marked accordingly and both to be put inside another envelop, that should be sealed and superscribed with tender notice no. and due date. The bidders may submit bids duly signed in their own letterheads.

Completed tender bids should reach the office of the **Registrar, Indian Association for the Cultivation of Science (IACS), 2A & 2B Raja S. C. Mullick Road, Jadavpur, Kolkata-700032** on or before the scheduled date and time specified below:

Tender Notice No.	NIT No. IACS/SPS/NP/2018-19/91, Date: 23-01-2019
Last date and time of submitting tender	21 Feb 2019 12 noon
Pre-bid meeting	08 Feb 2018 3PM (Room no 123A)
Date and time of opening tender	21 Feb 2019 3 PM
Place of opening tender	Room no 123A, IACS
Date and time of opening of Price Bid	The Price Bids of the bidders qualifying the technical bid will only be opened, the date of which will be intimated to the short-listed bidders at their email addresses. The rest of the bids will be rejected.
Contact	Email: camnp@iacs.res.in Tel.+91 33 2473 4971 [Ext 1733]

The technical bids will be opened first to judge/evaluate the technical specifications of the said instrument and thereafter the price bids of only technically qualified bidders will be opened.

Technical Bid Evaluation: The Technical Bids will be evaluated in the presence of the representatives of intending bidders who will be able to clarify technical aspects of their bids, if any, required by the Technical Evaluation Team.

Opening of price-bid: The Price Bids of the bidders qualifying the technical bid will only be opened, the date of which will be intimated to the short-listed bidders at their email addresses. The rest of the bids will be rejected.

Please note that IACS will not provide any accommodation or reimburse any expenses to any of the bidders for attending opening of technical bid.

Quotations received incomplete or beyond the stipulated time will be summarily rejected.

Bidders should submit their past experience for supplying and successful installation of similar units to other research Institutes/Universities/other organizations in India and abroad. Please provide documentary proofs of such successful installation and supportive documents that the instruments are running successfully.

1. TECHNICAL BID

The technical bid should contain technical specifications and should be kept in a separate sealed envelope duly super scribed as 'TECHNICAL BID' on the outer cover of the envelop as already detailed above. It should be clearly mentioned on the envelope as "Technical Specification for **Mobile Optical Spectrophoto/Fluorometer**". However, the tender is divided into part A and Part B. If only one part is quoted, that should be marked in the envelope.

Technical Specifications

A: Mobile Absorption spectrophotometer

1. Dual beam UV-Vis spectrophotometer with window based PC control
2. Light Source - XENON LAMP 75 KWatt or higher
3. SCAN speed – The scan speed should be high (>20000nm/min preferable)
4. Czerny-Turner Monochromator
5. Wave length Range: 190nm-1000 nm
6. 1.5 nm or better spectral bandwidth
7. The system should allow open air and room light measurement.
8. Fiber Optics connectivity for measuring sample externally.
9. Fiber Optics Probe (10 mm pathlength)
10. Computer Core i5, 13.3 inch touch screen laptop (Make Dell or HP)
11. Warranty – 1 year (minimum) (Extendable to 3 years). For 1 year warranty, two additional year AMC to be included.

Note: Entire system would be in a mobile rack and there should be IN and OUT of fiber optics cables. Items available in India may be quoted in INR.

B. Specification for Mobile Spectrofluorometer:

Detector

CCD array type detector, Range : 350-1000 nm, Pixels: 2048 pixels or better

Integration Time : 1 ms – 65 seconds, Dynamic Range: 1300:1 for a single acquisition or better

Resolution: 1.5 nm with 25um slit, Signal-to-noise ratio: 250:1

Detector collection lens should be installed

Interchangeable slit: Interchangeable slit; 5um; 10um; 25um; 50um; 100um and 200um,

Light Source

High power 265 nm, 365nm & 470nm LED light source

(a) Wavelength: 265nm

Output power: min 0.18mW or better

Maximum current: 350 mA or better

(b) Wavelength: 365 nm

Output power: min 10mW or better

Maximum current: 1200 mA or better

(c) Wavelength: 470 nm

Output power: min 3mW or better

Maximum current: 350 mA or better

For all, Current can be control via controller. The exact wavelength might vary and approval might be obtained for the same.

Software:

Real-time spectral acquisition and analysis software.

Accessories

Fiber: 600um Premium Fiber, 2m long – 2Nos

Cuvette Holder: Cuvette Holder, 1-cm path, for fluorescence

Collimating lens – 1no

Specification: Diameter: 25mm, Focal Length: 100mm, Wavelength range: 200-2000nm

Connector: SMA

Cuvettes: screw cap quartz cuvettes 10 mm pathlength – 8 nos(7 to 10 ml), 2 nos 3.5 ml

Computer:

System configuration: Core- i5 Processor, 13.3 to 14 inch screen portable laptop (Dell or HP)

Warranty: 1 Year (Extendable to three years) plus two years AMC

Other Services and Installation

Pre-installation requirements should be mentioned and intimated in advance.

Installation : This would be carried out by the company engineers with no extra cost.

Power supplies must conform to the mains supply: 230 $\pm$ 10 volts, 50 Hz.

Servicing facility within India is a must while service engineer stationed at Kolkata will be preferred.

A list of other places where the instrument has been installed should also be provided.

A compliance table (see below) must be prepared and submitted along with the technical bid.

Sr. No	Tender specification	Your offered instrument specification	Extent of compliance

2. PRICE BID

The financial bid indicating (item-wise) price for the item(s) mentioned in the technical bid should be kept in a separate sealed envelope duly super scribed as 'PRICE BID' on the outer cover of the envelop as already detailed above. Price bids of only technically qualified bidders will be opened and they will be intimated the date and time of the opening of price bid at their e-mail ids. Rest of the bidders will stand rejected.

PRICE: Price to be quoted on CIF Kolkata and also FOB basis.

3. BID SECURITY:

a. An Account payee Demand Draft of any of the Commercial Banks for Rs. 40,000 (Rupees forty thousand only) (For one part A or B, Rs 20,000/-) drawn in favour of "Indian Association for the Cultivation of Science (State Bank of India, Jadavpur University Branch, A/C No. 11079699211, IFSC: SBIN000093, MICR Code: 700002048)" is to be furnished by the bidders alongwith Technical Bid, except Micro and Small Enterprises (MSEs) as defined in MSE Procurement Policy issued by Department of Micro, Small and Medium

Enterprises (MSME) or are registered with the Central Purchase Organisation or the concerned Ministry or Department.

Alternatively, the bidder shall have the option to furnish EMD in the form of Fixed Deposit Receipt, Banker's Cheque or Bank Guarantee from any of the Commercial Banks. The Bid-Security should have at least 90 (ninety) days validity period beyond the final bid validity period.

In case of non-award of the work the Bid Security money would be returned to the unsuccessful Bidders.

4. PERFORMANCE SECURITY:

An Account Payee Demand Draft on any commercial bank of India or Bank Guarantee (BG) of any Commercial Bank preferably located at Kolkata of 10% of the order value in the name of “Indian Association for the Cultivation of Science” is to be furnished by the successful bidder as Performance security. Performance security money should remain valid for a period of 60 days beyond the date of completion of all contractual obligations of the supplier including warranty obligations. Bid security money or EMD will be refunded to successful bidder on receipt of the Performance security money.

5. TERMS OF PAYMENT:

Payment will be made through irrevocable Letter of Credit in two installments. 90% of the money will be released on submission of shipping of documents. Remaining 10% will be released after successful installation of the instrument.

6. GENERAL INSTRUCTIONS :

1. Validity of tender: Tender submitted should remain valid for at least six months from the date of opening the tender. Validity beyond six months from the date of opening of the tender shall be lapsed by mutual consent.
2. The tender should accompany a compliance chart.
3. Incomplete and conditional tenders as well as tenders received after the due date will be summarily rejected without assigning any reasons thereof.
4. At any time prior to the bid due date, IACS may, for any reason, whether at its own initiative or in response to a clarification requested by a prospective bidder during pre-bid meeting, modify the bidding documents. The amendment(s) will be notified on the institute website. Prospective bidders are advised to occasionally to visit the website (www.iacs.res.in/tender) for any amendment.
5. Installation/Demonstration/Application training at site: Installation & user training at IACS, free of cost by the supplier.
6. Service facility: In India, preferably Kolkata, supplier should mention their details of service setup and man powers who are responsible for after sales support. Response time should be within 24 hrs.
7. The model number, make and a printed literature of the product should be submitted positively.
8. Proposed delivery schedule should be mentioned clearly.
9. Manufacturers / exclusive distributors / vendors should have history of supplying this type of instruments to this or other scientific organizations. Availability of a list in this regard would be preferred.
10. Authorized dealership certificate should be provided in case of principal manufacturing company is not quoting directly.
11. Guarantee certificate, users manuals etc. are to be handed over to the user after successful commissioning of the system.
12. In the event of date being declared a closed holiday for purchaser's office, the due date for submission of bids and opening of the technical bids will be the following working day at the appointed time.

13. In case of any dispute, the decision of IACS authority shall be final and binding on the bidders.
14. In case of any litigation the matter will be dealt within the jurisdiction of the High Court of Calcutta only.
15. For any clarification regarding technical specifications, information etc., please send your queries to Narayan Pradhan (camnp@iacs.res.in).
16. The authority of IACS reserves the right to reject any or all of the tenders received without assigning any reason thereof.

Acting Registrar