

No. IACS/ CSS / SR&SM / 2016-17 / 73

Date: - 27.01.2017

CORRIGENDUM NOTICE No. II

Change in the specifications

Ref: NIT No. IACS/ CSS / SR&SM / 2016-17 / 73, Date: -03.01.2017

As decided in the technical committee meeting held on January 27, 2017, a correction is being made to the above mentioned tender notification and also the last date of submission of tender application as well as the date of opening the technical bid and price bid is being postponed to the following dates.

Correction:

Old specification: X-ray sources: Rotating/stationary with power 3 kW or higher.

New specification: X-ray generator power: Rotating anode/stationary tube source with x-ray generator power minimum of 3 kW or preferably more.

Modified dates:

Last date of submission of tender: February 10, 2017, 12 noon.

Date of opening the technical bids followed by the price bid: February 10, 2017, 4.00 pm.



Acting Registrar