

Indian Association for the Cultivation of Science
2A & 2B Raja S.C. Mullick Road, Jadavpur, Kolkata - 700032

Corrigendum of Tender

NIT No. : IACS/PSU/AKN/ 2017-18/83, Date: 02/01/2018

Tender Time Extension

Tender Notice No.	NIT No.: IACS/PSU/AKN/2017-18/83, Date: 02-01-18
Last date and time of submitting tender	31/01/2018 within 11.30 AM
Date and time of opening tender (Technical Bid)	31/01/2018 at 12.00 noon

Technical Specification

Corrigendum of the following points given in the Technical specifications

4. Rate of evaporation: **1200mL-7500mL/hr** (maximum) (acetone)
5. Power supply is **1400W or less due** to the saving of power.
6. Condenser surface area (cooling area): **1200-2000 cm²**
8. Not required. Remove it.
10. Heating bath capacity id within the **range of 170-200 °C**.

- 12, 13. Point No. 12 and 13 of the last e-tender, are not necessary and these points may be kept optional.
16. Sample flask capacity: **50ml-2000ml (max.)**,
flask- 2 in numbers and 500 ml flask- 2 in number are required.

20. Vacuum pump
 - a) Teflon made diaphragm vacuum pump with specification **5-10 mBar**.
 - b) Aspirator or Teflon made diaphragm pump with **10±5 mBar**
(Pump should be Teflon made).

21. Vacuum controller: Digital controller with digital display programmable mode preferable.
24. All components from the same manufacturer
26. Period of warranty: 2 years.

Acting Registrar