

INDIAN ASSOCIATION FOR THE CULTIVATION OF SCIENCE
2A & 2B Raja S.C. Mullick Road, Jadavpur, Kolkata - 700032

Ref. No.- SMS/AB/19-20/08-1

DATE : 13-05-2019

CORRIGENDUM NOTICE No.- I

NIT No. - IACS/SMS/AB/2019-20/08 DATE : 25.04.2019

Technical Specifications:-

1. Page No.2, Item No. F, 'Flameproof Design heater' should read as 'Heater'
2. Page No.2, Item No. H, 'variable frequency drive' should read as 'variable frequency drive or tachometer assembly'
3. Page No.2, Item No. K, 'Pressure gauge with vacuum protection' should read as 'Pressure gauge with vacuum protection/valve for pressure release'
4. Page No.3, Item No. N, should read as 'Digital temperature indicator should be placed near the autoclave so that control panel can be remotely monitored at upto 5 m in a safe area'
5. Page No.3, Item No. W, 'SS movable trolley with all 4 lockable wheels' should read as 'SS movable trolley with all 4 lockable wheels or integrated table top model without trolley.

ACTING REGISTRAR